

Instrument SerialNumber: 010523 **Test #:** 001409 **Subject Test**
Start Date/Time: 10/17/2020 03:41 **System Check:** Passed **Remote/Local:** Local
Operator's Last Name: PHAROAH **First Name:** IAN **Middle Initial:** M
Agency: Metropolitan Airport Aut **License Number:** 31247
Card Serial Number: 131247 **Effective Date:** 09/01/2019 **Expiration Date:** 09/01/2021
Court Name: FAIRFAX COUNTY GDC **End Date/Time:** 10/17/2020 03:49
Reported AC Sample Date/Time: 10/17/2020 03:46 **Reported AC:** 0.06
Data Type: DIAG **Sample Value:** Pass **Sample Time:** 03:41
Data Type: BLK **Sample Value:** 0.000 **Sample Time:** 03:42
Data Type: CHK **Sample Value:** 0.099 **Sample Time:** 03:43
Data Type: BLK **Sample Value:** 0.000 **Sample Time:** 03:44
Data Type: SUBJ **Sample Value:** 0.062 **Sample Time:** 03:46
Data Type: BLK **Sample Value:** 0.000 **Sample Time:** 03:47
Data Type: SUBJ **Sample Value:** 0.063 **Sample Time:** 03:48
Data Type: BLK **Sample Value:** 0.000 **Sample Time:** 03:49
Standard Type: Dry Gas Std **Standard Value:** 0.100 **Standard Lot Number:** AG908501-07
Standard Expiration Date: 03/26/2021 **Tank Pressure:** 860 **Barometric Pressure:** 757 mmHg
Blow Sample Number: 1 **Blow Duration:** 0.58 sec **Blow Volume:** 126 cc **End of Blow Time:** 03:45
Blow Sample Number: 2 **Blow Duration:** 4.08 sec **Blow Volume:** 1579 cc **End of Blow Time:** 03:46
Blow Sample Number: 3 **Blow Duration:** 5.81 sec **Blow Volume:** 2142 cc **End of Blow Time:** 03:48
Test Status: Success **Test Status Code:** 0 **Tamper Evident Stamp:** 40bd3102

Instrument SerialNumber: 010523 **Test #:** 001410 **Subject Test**
Start Date/Time: 10/17/2020 06:06 **System Check:** Passed **Remote/Local:** Local
Operator's Last Name: WHISENHUNT **First Name:** RICHARD **Middle Initial:** M
Agency: Fairfax County PD **License Number:** 31765
Card Serial Number: 131765 **Effective Date:** 06/01/2020 **Expiration Date:** 06/01/2022
Court Name: FAIRFAX COUNTY GEN DIST **End Date/Time:** 10/17/2020 06:13
Reported AC Sample Date/Time: 10/17/2020 06:12 **Reported AC:**
Data Type: DIAG **Sample Value:** Pass **Sample Time:** 06:06
Data Type: BLK **Sample Value:** 0.000 **Sample Time:** 06:07
Data Type: CHK **Sample Value:** 0.099 **Sample Time:** 06:07
Data Type: BLK **Sample Value:** 0.000 **Sample Time:** 06:09
Data Type: SUBJ **Sample Value:** * ** **Sample Time:** 06:12
Data Type: BLK **Sample Value:** 0.000 **Sample Time:** 06:13
Standard Type: Dry Gas Std **Standard Value:** 0.100 **Standard Lot Number:** AG908501-07
Standard Expiration Date: 03/26/2021 **Tank Pressure:** 860 **Barometric Pressure:** 758 mmHg
Blow Sample Number: 1 **Blow Duration:** 0.32 sec **Blow Volume:** 70 cc **End of Blow Time:** 06:10
Blow Sample Number: 2 **Blow Duration:** 0.90 sec **Blow Volume:** 205 cc **End of Blow Time:** 06:12
Test Status: Deficient Sample **Test Status Code:** 45 **Tamper Evident Stamp:** 253d9699

Instrument SerialNumber: 010523 **Test #:** 001411 **Subject Test**
Start Date/Time: 10/17/2020 06:14 **System Check:** Passed **Remote/Local:** Local
Operator's Last Name: WHISENHUNT **First Name:** RICHARD **Middle Initial:** M
Agency: Fairfax County PD **License Number:** 31765
Card Serial Number: 131765 **Effective Date:** 06/01/2020 **Expiration Date:** 06/01/2022
Court Name: FAIRFAX COUNTY GEN DIST **End Date/Time:** 10/17/2020 06:25
Reported AC Sample Date/Time: 10/17/2020 06:21 **Reported AC:** 0.13
Data Type: DIAG **Sample Value:** Pass **Sample Time:** 06:14
Data Type: BLK **Sample Value:** 0.000 **Sample Time:** 06:15
Data Type: CHK **Sample Value:** 0.099 **Sample Time:** 06:16
Data Type: BLK **Sample Value:** 0.000 **Sample Time:** 06:17
Data Type: SUBJ **Sample Value:** 0.137 **Sample Time:** 06:21
Data Type: BLK **Sample Value:** 0.000 **Sample Time:** 06:22
Data Type: SUBJ **Sample Value:** 0.153 **Sample Time:** 06:23
Data Type: BLK **Sample Value:** 0.000 **Sample Time:** 06:25
Standard Type: Dry Gas Std **Standard Value:** 0.100 **Standard Lot Number:** AG908501-07
Standard Expiration Date: 03/26/2021 **Tank Pressure:** 855 **Barometric Pressure:** 758 mmHg
Blow Sample Number: 1 **Blow Duration:** 2.79 sec **Blow Volume:** 691 cc **End of Blow Time:** 06:18
Blow Sample Number: 2 **Blow Duration:** 3.41 sec **Blow Volume:** 1326 cc **End of Blow Time:** 06:19
Blow Sample Number: 3 **Blow Duration:** 4.09 sec **Blow Volume:** 1604 cc **End of Blow Time:** 06:21
Blow Sample Number: 4 **Blow Duration:** 4.78 sec **Blow Volume:** 1995 cc **End of Blow Time:** 06:23
Test Status: Success **Test Status Code:** 0 **Tamper Evident Stamp:** 998fa602

